The analysis application of degraded natural rubber EI-009 by ionRocket-DART[®]-MS

Natural rubber (NR), Deterioration, Additive, Life prediction, ionRocket, DART® [Keyword] [Subject area] Material analysis, Quality control (QC), Polymer manufacturing

Abstract

Natural rubber (isoprene polymers and minor impurities), is prone to rapid degradation. To protect natural rubber from chemical and physical attacks, many chemicals are added during coagulation, processing and molding. The combination and amount of chemicals play an important role in the life time of rubber products. Consequently, methods used to estimate the degree of degradation such as FT-IR and now DART[®]-MS are critical for both development and OC operations in rubber manufacturing.

DART[®]-MS (Direct Analysis in Real Time - Mass Spectroscopy) allows direct ionization of solid and liquid samples at atmospheric pressure, obviating almost all sample preparation steps. ionRocket improves DART[®] for polymers by subjecting the sample to a rapid temperature gradient. By using ionRocket, both small molecules and high molecular weight, difficult to volatilize compounds can be detected in the same run. In this report, ionRocket-DART[®]-MS is applied to a degraded rubber band and the results compared with FT-IR analysis.

Sample

Rubber band (commercial item)


Method

The rubber band was heat treated at 130 °C for 30 min. Sample preparation: the rubber band was sliced into 0.5 mm x 0.5 mm sections, and then placed in the ionRocket copper sample pot. A temperature gradient of 100 °C/min. from room temperature to 600 °C was applied during a total run time of 7 min.

Result

The result of FT-IR by ATR (Attenuated Total Reflection) method is shown in Figure 1. In general, C=C bonds of main chains of natural rubber are decreased and >C=O is increased by heat degradation. However, FT-IR did not show any distinct differences.

ionRocket-DART[®]-MS analysis results are shown in Figures 2 and 3, which show distinct differences due to the heat treatment. The intensity of m/z 136 at 5 min (about 400 °C) was decreased after treatment. This peak was probably from additives. In addition, the intensity of m/z 373 was increased after treatment. This peak was probably from rubber degradants.









Figure 2. *m/z* vs time-temperature vs intensity of new rubber band



vs intensity of degraded rubber band



